



नेपाल सरकार  
गृह मन्त्रालय



## मध्य पश्चिम क्षेत्रीय प्रहरी अस्पताल

(क्याटलग सपिड) विधिबाट सवारी साधन (एम्बुलेन्स) खरिद गर्ने सम्बन्धी सूचना

सूचना नं. ०१/२०७७/२०७८

प्रथम पटक सूचना प्रकाशित मिति २०७७।१२।२९ गते ।

मध्य पश्चिम क्षेत्रीय प्रहरी अस्पताल नेपालगंजलाई आवश्यक पर्ने सवारी साधन (आई.सि.यु जडान सहितको क बर्गको एम्बुलेन्स) खरिद गर्नु पर्ने भएकोले उत्पादक वा अधिकृत बिक्रेताद्वारा निर्धारित दरमा (क्याटलग सपिड) खरिद गर्ने कार्यविधि २०७४, सार्वजनिक खरिद ऐन २०६३ (संशोधन सहित)को दफा ८ को उपदफा १ (क) तथा सार्वजनिक खरिद नियमावली, २०६४ (संशोधन सहित)को नियम ३१ (ख) बमोजिम त्यस्तो सवारी साधन उत्पादक कम्पनी वा सो को आधिकारीक बिक्रेताहरू बीच मात्र प्रतिस्पर्धा गराउने (क्याटलग सपिड) विधिबाट खरिद गर्नु पर्ने भएकोले ईच्छुक इजाजत प्राप्त उत्पादक कम्पनी वा त्यसको आधिकारीक बिक्रेताहरूले आ-आफ्नो फर्म दर्ता, एजेन्सी दर्ता प्रमाणपत्र Letter of Authorization र आ.व. २०७६/२०७७ सालको कर चुक्ता प्रमाण पत्रको प्रमाणित प्रतिलिपी (नोटरी पब्लिक) प्रतिलिपहरू समावेश गरी सार्वजनिक खरिद नियमावली २०६४ (संशोधन सहित) को नियम ३१ ख (१) अनुसार देहाय बमोजिमको सवारी साधनको उत्पादनको आधिकारीक स्पेशिफिकेशन, मुल्य र सुविधा सहितको गुणस्तर विवरण (क्याटलग वा ब्रोसर) संग्लन राखी यो सूचना प्रकाशित भएको मितिले ८ औं दिन मिति २०७८।०१।०८ गते १२:०० बजे भित्र यस कार्यालयको सिलबन्दी प्रस्ताव दर्ता गर्नु हुन सम्बन्धित उत्पादक वा आधिकारीक बिक्रेताको जानकारी लागी यो सूचना प्रकाशित गरिएको छ ।

अस्पताल प्रमुख  
(डा. अभिमन्यु श्रेष्ठ)  
प्रा. प्रहरी उपरीक्षक


सि.न.	सवारी साधन	संक्षिप्त विवरण	कैफियत
१.	ICU Equipped Ambulance Van "A" Category	The RHD ICU Equipped Ambulance Van, must operate with in the temperature range (-5 degree centigrade to +55 degree centigrade) and the altitude up to 2000 meter, suitable for rough road operation in Nepal. Seating capacity Minimum 4 Persons .Must have Advance Life Support medical Equipment, Category "A" should keep in the Ambulance With appropriate arrangement. 4 Stroke Diesel Engine Minimum 4 Cylinders. Piston Displacement min-2100 cc Engine Power.55kw@ rated RPM Engine Torque –min 200Nm@rated RPM Colour –As per Standard Norms. Transmission Minimum 5 speed all synchromesh gear . Warranty a Minimum one year after delivery	संक्षिप्त विवरणमा उल्लेख नभएका कुराहरू स्वीकृत Technical Specification अनुसार हुने छ । यो सूचना लुम्बिनी प्रदेश प्रहरी कार्यालयको Website मा समेत राखीएको छ ।

पुनश्च :- प्रबिधिक प्रस्ताव पेश गर्दा यस अस्पतालबाट उपलब्ध हुने Technical Specification लिई Offered Specification महलमा भरी पेश गर्नु पर्ने छ । साथै थप जानकारीको लागि यस अस्पतालको आ.प्र. शाखाको ९८४९७८५४४० मा सम्पर्क गर्न हुन अनुरोध गरिएको छ ।

### Specification of Ambulance Van

S.N.	Criteria	Requirements	Offer Required
1	General	The RHD ICU Equipped Ambulance Van, must operate with in the temperature range (-5 degree centigrade to +55 degree centigrade) and the altitude up to 2000 meter, suitable for rough road operation in Nepal. Seating capacity minimum 4 persons .	
2	Engine	4 Stroke Diesel engine. Cylinders: Min. 4. Piston Displacement capacity: Min. 2100cc. Water cooled . Power Min: 55KW @ rated RPM. Torque Min : 200 Nm @rated RPM .	
3	Gauges , Meter Indicator	Fog Lights set. OEM should be supplied. Oil pressure warning indicator. Accessory meter, indicator gauge & warning buzzer will be highly appreciated.	
4	Transmission	Clutch type - Single dry disc plate. Min. 5 speed all synchromesh gear.	
5	Brake	Hydraulic double circuit brake system with power assisted. (tandem master cylinder ). Front: - Ventilated Disc. Rear: - Drum or Disc. Suitable Parking brake . Anti Lock Braking System.	
6	Steering system	Power steering .	
7	Tyre	STD Size Tyre. Extra Spare wheel with same Standard.	
8	Suspension	<b>Front:</b> McPherson struts with coil spring type, or Semi elliptical leaf spring with shock absorber. <b>Rear:</b> Parabolic leaf spring or semi elliptical leaf spring with shock absorbers.	
9	Colour	As per Standard Norms.	
10	Essential Accessories	Multiflashing Beacon Bar Lamp (Red). Powerful siren min. 3 stage sound. Place for communication radio set with clamp. Cigarette lighter. Side Mirror. Air Condition system. CD, MP3 /USB FM, AM Radio player. Sufficient internal lighting . Rubber Mats. Fire Extinguisher min. 1 kg. Hydraulic jack with essential tools	
11	Dimension	Wheel Base: Minimum 2900 mm Ground Clearance: Min. 170 mm	
12	Weight & Capabilities	Fuel Tank Capacity: Min. 60 Liters GVW: Min. 2900 Kg	
13	Electrical	Battery: 12 volts Electrical Control System.	

*[Handwritten signatures]*



14	Warranty	Manufacturer shall provide a minimum of one year warranty after delivery. Supplier must provide training for at least 2 persons for the Medical equipment handling.	
15	Emission Control	The Emission Standard of the vehicle should be capable of meeting applicable Nepal Vehicle Mass Emission Standard 2069 which must be certified by Department of Transport Management, Nepal.	
16	Manual	Owner's instruction book per unit. All signs and instruction in the equipment shall be in English.	
17	Delivery	As per Contract.	

#### **Ambulance Equipment**

18	Medical Equipments	Must have Advance Life Support medical Equipment, <b>category 'A'</b> should keep in the Ambulance with appropriate arrangement.	
		<b>All Emergency Equipments</b>	
		Stethoscope	
		B.P. Apparatus	
		Torch Light	
		Tongue Depressor	
		IV Drips	
		Cannulas and Syringes	
		ECG Monitor with oxygen Monitor	
		Intubation Set	
		Various Intubation Tubes and Laryngeal Tubes	
		Ambu bag	
		Travelling Ventilator	
		Nebulizer set	
		Manual Suction set	
		Cervical Collars	
		CPR Board	
		Oxygen Supply	
		Automated external defibrillator (AED)	
		OB kit	
Slpints			
Catheterizations Sets			
Chest Drainage Tubes			
Haemostatic Sets			
19	Others	Washing Equipments.	
		Wheel chair and trolley	
		Radio Communication	
		Tablet with 4G Connectivity.	

**Prepared By:**

Insp. Anup Shrestha

Insp. Er. Anjan Raj Kharel



**Approved By**

SSP Er. Narottam Lal Pradhan

PHQ, MT Division, Naxal, Kathmandu

Date: 2077/12/17